

See discussions, stats, and author profiles for this publication at: <https://www.researchgate.net/publication/371226439>

The Conventional Diagnostic Techniques of Malaria in Endemic Areas: A Review of the Approaches with Focus on Newly Noninvasive Methods

Article · June 2023

DOI: 10.31031/CJMI.2023.06.000645

CITATIONS

0

READS

43

3 authors:



Ammar Abdelmola
Jazan University

7 PUBLICATIONS 0 CITATIONS

[SEE PROFILE](#)



Ghanem Mohammed Mahjaf
Shendi University

59 PUBLICATIONS 3 CITATIONS

[SEE PROFILE](#)



Mosab Nouraldein Mohammed Hamad
Elsheikh Abdallah Elbadri University

552 PUBLICATIONS 156 CITATIONS

[SEE PROFILE](#)

Some of the authors of this publication are also working on these related projects:



Researchers Academy Excellence [View project](#)



teaching materials [View project](#)

The Conventional Diagnostic Techniques of Malaria in Endemic Areas: A Review of the Approaches with Focus on Newly Noninvasive Methods

ISSN: 2578-0190



***Corresponding author:** Mosab Nouraldein Mohammed Hamad, Assistant professor, Microbiology department, Faculty of Medicine, Elsheikh Abdallah Elbadri University, Sudan

Submission:  March 17, 2023

Published:  May 26, 2023

Volume 6 - Issue 4

How to cite this article: Ammar A Abdelmola, Ghanem Mohammed Mahjaf and Mosab Nouraldein Mohammed Hamad*. The Conventional Diagnostic Techniques of Malaria in Endemic Areas: A Review of the Approaches with Focus on Newly Noninvasive Methods. *Cohesive J Microbiol Infect Dis.* 6(4). CJMI. 000645. 2023.

DOI: [10.31031/CJMI.2023.06.000645](https://doi.org/10.31031/CJMI.2023.06.000645)

Copyright@ Nouraldein Mohammed Hamad. This article is distributed under the terms of the Creative Commons Attribution 4.0 International License, which permits unrestricted use and redistribution provided that the original author and source are credited.

Ammar A Abdelmola¹, Ghanem Mohammed Mahjaf² and Mosab Nouraldein Mohammed Hamad^{3*}

¹Department of Medical Laboratory Technology, Jazan University, Saudi Arabia

²Department of Medical Microbiology, Shendi University, Sudan

³Assistant professor; Microbiology Department, Elsheikh Abdallah Elbadri University, Sudan

Abstract

Background: Malaria continues to be a major global health problem, with over 228 million cases and 405,000 deaths estimated to occur annually. Rapid and accurate diagnosis of malaria is essential to decrease the burden and impact of this disease, particularly in children. We aimed to review the main available techniques for the diagnosis of clinical malaria in endemic settings and explore possible future options to improve its rapid recognition.

Methods: Electronic databases including, Google, Google Scholar, PMC, PubMed, Science Direct, and Scopus were rigorously searched using the terms Malaria diagnosis, Microscopy for malaria, plasmodium detection, Rapid diagnostic tests, PCR for malaria for the completion of this descriptive review.

Result: Literature review shows that Light microscopy is still considered the gold standard method for malaria diagnosis and continues to be at the frontline of malaria diagnosis. However, technologies such as rapid diagnostic tests, mainly those that detect histidine-rich protein-2, offer an accurate, fast, and affordable alternative for malaria diagnosis in endemic areas. They are now the technique most extended in endemic areas for parasitological confirmation. In these settings, PCR-based assays are usually restricted to research and they are not currently helpful in the management of clinical malaria. Other technologies, such as isothermal methods could be an exciting and alternative approach to PCR in the future.

Conclusion: Available evidence suggests that the role of RDT, despite its increasing false negatives, is still the most feasible diagnostic test because it is easy to use, fast and does not need expensive equipment. Noninvasive tests that do not require a blood sample, but use saliva or urine, are some of the recent tests under development that have the potential to aid malaria control and elimination. Therefore, future innovation will be required to apply more sensitive and affordable methods in resource-limited settings.

Keywords: Malaria; Plasmodium; Rapid Diagnostic Test (RDT); Polymerase Chain Reaction (PCR); Loop-mediated Isothermal Amplification (LAMP)

Introduction

For centuries, humans have been plagued with malaria, a disease that seems to prevail over strategies used to combat it. It is becoming more challenging through the emergence of antimalarial drug resistance. The female Anopheles mosquito serves as a competent vector to transmit the Plasmodium parasite to human hosts with each blood meal. The mosquito remains unharmed by the parasite, and its ubiquitous nature ensures the transmissibility of the disease. According to the World Health Organization (WHO), there were 228 million cases and 405,000 deaths worldwide in 2018 [1]. The countries where the disease is endemic are referred to as the "malaria belt". These locations mostly have tropical climates that are conducive to the breeding of mosquitoes and subsequent transmission of the parasite to human hosts. However, several countries that were in this category have managed to

eliminate the disease and have been declared “malaria-free” by the WHO, due to a combination of programmatic approaches that involve early diagnosis and treatment [2]. The novel malaria RTS, S subunit vaccine seems to be a promising start for the use of vaccination as a strategy towards malaria elimination, however, with a vaccine efficacy of only 30% to 50%, there is still a long way to go before vaccination can be considered a reliable method against malaria [3]. Until such a time, current control measures need to be improved and deployed to their maximum capacity. Four well-established species of the malaria parasite infect humans, namely *Plasmodium falciparum*, *P. vivax*, *P. ovale*, and *P. malariae*. *P. falciparum*, the deadliest species, accounts for 99.7% of infections in Sub-Saharan Africa. *P. vivax* is the most common in the Americas and accounts for 75% of infections [1]. For Asia and Oceania, the number of *P. falciparum* and *P. vivax* infections are relatively equivalent. *P. ovale* and *P. malariae* are widely dispersed but have low incidence [4]. An additional species, *P. knowlesi*, is a simian malaria parasite that is usually found in long-tailed and pig-tailed macaques. However, zoonotic human infections have been reported in Southeast Asia [5]. Malaria treatment success relies on prompt diagnosis and recommendations of the most appropriate treatment regimen. Artemisinin-based Combination Therapy (ACT) is recommended by the WHO for the treatment of uncomplicated *P. falciparum* malaria and ACT or chloroquine is recommended for *P. vivax*, *P. ovale*, *P. malariae* and *P. knowlesi* malaria infections. Primaquine is recommended to prevent the relapse of *P. vivax* and *P. ovale* infections. These recommendations are modified for special groups such as pregnant or lactating women, patients with other comorbidities, individuals with Glucose-6-Phosphate Dehydrogenase (G6PD) deficiency individuals and young children and infants [6]. Malaria is a febrile illness and clinical symptoms of uncomplicated malaria include fatigue, headaches, muscle aches, malaise, abdominal discomfort, fever, nausea and vomiting [7]. Specific diagnostic methods are needed to differentiate between malaria and other febrile illnesses. An early diagnosis of malaria can prevent further progression and lower the severity of the disease. This is especially critical for children under five years of age who accounted for about 67% of the deaths in 2018 due to severe malaria worldwide [1]. For the most effective treatment of malaria, it is important to know the species of Plasmodium a person is infected with and the parasitic burden in the blood. Accurate, prompt and affordable diagnostic tools are also vital for tracking successes or drawbacks of control and elimination efforts, and for future programs aimed at global malaria eradication. Active surveillance of disease in each geographical area is essential for a program to succeed. The WHO Global Technical Strategy for Malaria aims, by 2030, to reduce malaria case incidence and mortality rates globally by 90%; to eliminate malaria from 35 countries in which malaria was transmitted in 2015; and to prevent the re-establishment of malaria in all countries that became malaria-free [8]. These targets, though ambitious, are important to set to challenge and remind the world that malaria is an important public health problem in need of serious and expanded efforts. Even with the existing challenges, several countries have succeeded in eliminating malaria. Countries

such as Algeria, Argentina, Paraguay, Sri Lanka and the Maldives have been declared malaria-free by WHO in recent years [2]. This shows that with appropriate programmatic approaches, malaria incidence can be substantially reduced. The goal of eradicating malaria by 2050 set by the Lancet commission [9] is an ambitious feat and may be possible only if all stakeholders are equally committed to the goal of eradication. The Lancet commission also emphasized the importance of ultrasensitive Rapid Diagnostic Tests (RDTs) as essential tools in the identification of asymptomatic infections and infections in pregnant women, and the need for novel diagnostic tools that do not require a finger-prick blood sample [9].

Current Malaria Diagnostic Options

Microscopy

The microscopic examination of thick and thin blood films is a “gold standard” test that is used to detect parasitemia in the blood and guide appropriate treatment. A drop of blood is collected from a patient via a finger stick or venipuncture. When venipuncture is used for blood collection, it is suggested that the blood is spread onto a slide immediately after collection to prevent prolonged exposure to anticoagulants in the collection tube that may alter parasite morphology [10]. Thick smears are more sensitive and involve placing one to two drops of blood on a slide in a circle. The red blood cells are lysed and the various malaria parasite blood-stages, trophozoites, gametocytes and schizonts are released. Thin smears are used to detect the morphology of the parasite species and are prepared by spreading a drop of blood across a slide to create a feathered edge that contains a single layer of cells [10]. The slide was stained with the Giemsa stain and examined using an Olympus bright-field microscope (BH-2, Tokyo, Japan) (100× oil immersion) by a trained laboratory personnel. The sensitivity and specificity for this method is 95% and 98%, respectively when Polymerase Chain Reaction (PCR) is used for comparison [11]. The limit of detection for this method is approximately 50-200 parasites per μL of blood [12,13]. A skilled laboratory personnel is able to quantify parasitemia in a blood smear in about 60 min. excluding the cost of labor and obtaining a microscope, each test costs approximately \$0.12-\$0.40 [13]. Advantages and Limitations a blood film examination under a microscope allows for the identification of parasitemia percentage, parasitic morphology, and speciation. This method requires trained personnel and sensitivity, and specificity may vary based on the skill of personnel. The time it takes for parasitemia detection and quantification is long and may lead to a delay in treatment. The limit of detection is also not ideal, because sub-microscopic asymptomatic individuals with low parasitemia remain undiagnosed and untreated, and also enables the transmission cycle to continue in the community.

Rapid Diagnostic Test (RDT)

RDTs are designed to detect antigens and they involve the use of an immunochromatographic strip where the blood is dropped into one end and the results are depicted by lines on the strip surface [14]. Three types of antigens have been employed in this method, Plasmodium histidine-rich protein (HRP) 2 (pHRP-2), Plasmodium

Lactate Dehydrogenase (pLDH), and *Plasmodium aldolase*. pHRP-2 is specific to *P. falciparum*, while pLDH and Plasmodium aldolase are found in all species [15]. More than 90% of commercially available RDTs target pHRP-2 [16]. Antibodies immobilized on the surface of the test strip detect these parasitic antigens when the blood and buffer mixture migrate across it. Each RDT contains a positive control to indicate the validity of the test. Currently available species-specific RDTs are only able to identify *P. falciparum* and *P. vivax* species [17]. For other species, the RDT is only able to indicate the presence of the parasite alone without speciation. The WHO, in collaboration with the Foundation for Innovative New Diagnostics (FIND) and the Centers for Disease Control and Prevention, and other partners, conducted a review of various commercially available RDT brands. The WHO selection criteria for procurement of RDTs are used to evaluate each brand, and performance is noted with a panel detection score [18]. The WHO recommends proactive and reactive post-market surveillance to ensure that each lot deployed for use is up to standard [19]. The sensitivity of RDTs ranges from 85% to 94.8% and the specificity ranges from 95.2% to 99% [13,20,21]. The limit of the detection is comparable to that of microscopy, 50-100 parasites per μL of blood, and trained personnel can produce results in 15 to 20 min from the time of blood collection [12]. RDTs are relatively affordable with each test costing about \$0.85 [11]. An ultra-sensitive rapid diagnostic test is currently under development and could have the potential for up to 10-fold better detection limit than that of currently used RDTs and can detect the incidence of the disease up to a day and a half sooner [15]. Advantages and Limitations RDTs represent a fast and affordable method for malaria diagnosis [7]. RDTs are easy to deploy in resource-limited and hard-to-reach settings. The personnel training required is much less intensive as compared to microscopy and PCR. Community health workers can perform the test in their communities and then prescribe treatment or refer patients to healthcare centers. False negatives are becoming more common due to parasites with pHRP-2 gene deletion and prozone phenomenon in patients with high parasitemia [14]. RDTs may be unable to keep up with the ever-evolving nature of the malaria parasite and its changing epidemiology. The RDT method does not allow for the quantification of parasitemia and consequently monitoring therapy effectiveness is difficult. It can also result in false positives because it detects pHRP-2 which can remain in the blood up to 30 days after treatment and effective elimination of active infection [14]. The limit of detection does not allow for the identification of asymptomatic carriers, and the variation in the performance of different RDT brands could lead to decreased reliability of the method.

Polymerase Chain Reaction (PCR)

PCR-based methods identify the presence of malaria target genes in a blood sample. There are various modifications of this test including, nested conventional PCR, multiplex real-time PCR and reverse transcriptase PCR [22]. Most of these methods target genes on the 18S rRNA of the malaria parasite [11]. PCR-based tests can be used for initial testing of suspected malaria cases and

parasite species; however, microscopy is often used to quantify parasitemia. PCR-based tests are uniquely useful to identify asymptomatic and submicroscopic patients that microscopy and RDTs miss. The sensitivity and specificity for the various PCR type's ranges from 98% to 100% and 88% to 94%, respectively when microscopy was used as the gold standard [23-25], and the limit of the detection is 0.5-5 parasites per μL of blood [11]. The test is typically completed in two hours, and apart from the initial cost of an expensive thermocycler and other equipment, each test costs \$7 to \$8 [21]. Advantages and Limitations The PCR-based methods are particularly useful in parasite detection in individuals with low parasite burden. The sensitivity and specificity are higher than that of RDTs and microscopy. PCR-based methods require the acquisition of a thermocycler which may be a financial hindrance for resource-limited settings to adopt the method. It also needs highly skilled personnel to perform the test and is not feasible for use in field settings. The PCR method also does not provide an easy method of estimating parasite burden that is often used by clinicians to make treatment decisions.

Novel Malaria Diagnostic Options under Development

Loop-Mediated Isothermal Amplification (LAMP)

LAMP is a relatively newer method for nucleic-acid amplification first described in 2000, and further modified for ease of visualization of amplified products using a fluorescent or colorimetric dye such as calcein and Hydroxy Naphthol Blue (HNB) respectively [25,26]. Unlike PCR, the LAMP procedure can be carried out in a 65 °C bath or a heat block for 30 to 60 min [21]. The sensitivity of LAMP ranges from 98.3% to 100% and specificity from 94.3% to 100% when compared to microscopy [26]. Commercially available test kits target the 18S rRNA of the malaria parasite, just like PCR [27]. LAMP performed with mitochondrial DNA targets has also been shown to have greater sensitivity and takes less time than that of 18S rRNA [27]. The limit of detection for this method is 1-5 parasites per μL of blood and the cost per test is estimated to be around \$5.31 [21]. Advantages and Limitations the limit of detection by LAMP is comparable to that of PCR because they are both in the range of 0.5-5 parasites per μL of blood. It is faster than PCR and the results can be assessed visually without the need for any expensive thermocycler. However, the method requires moderately skilled personnel and has a complex primer design.

Nucleic Acid Sequence-Based Amplification (NASBA)

NASBA is a diagnostic method that involves the use of three enzymes, reverse transcriptase, T7 RNA polymerase, and RNase H, to amplify RNA targets in a double-stranded DNA background [21]. The RNA target, such as 18S RNA, is copied into complementary DNA (cDNA) using reverse transcriptase and then the cDNA is amplified using T7 RNA polymerase [11]. It does not require the use of a thermocycler because the reaction can be carried out at 41 °C resulting in more than 108-fold amplification of the target RNA sequence. The sensitivity of the method when compared to

microscopy ranges from 97.40% to 100% while the specificity ranges from 80.90% to 94% [21]. The limit of detection is 0.01-0.1 parasites per μL of blood. The test is estimated to take about one hour to complete and costs between \$5 and \$20 per test [11]. Advantages and Limitations NASBA, like LAMP, does not require a thermocycler and has a very low limit of detection. However, it requires extensive training of personnel to ensure the reliability of the results and the cost per test is much higher than other methods.

Isothermal Thermophilic Helicase-Dependent Amplification (tHDA)

In tHDA, the double-stranded DNA is separated by helicase and single-stranded DNA-binding proteins are attached to the separated strands. Specific primers bind to the strands and DNA polymerase synthesizes new strands, and the test is performed at 65 °C for about two hours [21]. In the application of tHDA for malaria diagnosis, the 18S rRNA gene is amplified from whole blood directly without heat denaturation or nucleic acid amplification. Probes labeled with either Fluorescein (FAM) or Digoxigenin (DIG) hybridize to the amplicon and the amplification product is detected with a lateral-flow strip that contains anti-FAM or anti-DIG antibodies. The sensitivity and specificity of this method are 96.6% and 100%, respectively when microscopy is used as the gold standard [28]. The limit of detection is 200 parasites per μL of blood and the results can be obtained in one to two hours [21,28]. Advantages and Limitations the tHDA method does not require the use of a thermocycler so the cost may be more affordable than PCR. In addition to its more affordable cost, whole blood can be used directly without any manipulation, hence simplifying the method. However, the limit of detection is higher than in any other nucleic acid-based methods. Although the method requires minimally trained personnel, the higher limit of detection is not suitable to detect malaria in patients with the low parasitic burden.

Saliva-based test with nucleic-acid amplification

This saliva-based malaria diagnosis involves the detection of a Plasmodium gene, 18S rRNA, or *P. falciparum* dihydrofolate reductase gene in saliva using a Nested-Polymerase Chain Reaction (n PCR). An n PCR involves the same procedure as conventional PCR but uses two primer sets and has two successive PCR steps. The product from the first PCR reaction serves as the template for the second reaction [29]. A thermocycler is required for the procedure, and nucleic acid has to be extracted first from the saliva sample before n PCR is performed. The sensitivity and specificity of this method range from 86.36% to 95%. The specificity ranges from 93% to 98.4% when compared with microscopy [29,30]. The limit of detection for this method is 1-10 parasites per μL of blood [31]. Advantages and Limitations Saliva-based tests are noninvasive and require less training for health personnel for sample collection. However, the method still involves the use of PCR, and health personnel will need advanced training on the actual PCR protocol. The procedure takes approximately six hours to complete, so this can be a major hindrance to its implementation. It can be estimated that the cost of the test will be similar to that of a blood-based PCR

test since it requires two PCR reactions using a thermocycler.

Saliva-based plasmodium protein detection

This saliva-based test detects the presence of specific Plasmodium proteins in the saliva of an infected person even before symptoms begin. A study conducted in Nigeria [32] used a commercially available kit, Opti MAL-IT dipstick (RDT), to detect p LDH in the saliva of children. The sensitivity of the test using whole saliva was 77.9% [32]. However, a study in Mali [33] reported a higher sensitivity of 97.2% and a specificity of 95.4% when compared to microscopy [33]. The limit of detection for this method is high at about 1000 parasites per μL of blood. In a separate study [34] conducted in Cameroon, Zambia, and Sierra Leone, a prototype Lateral-Flow Immuno Assay (LFIA) was developed to detect the presence of PSSP17 protein, a female gametocyte protein that is presumably more abundant in saliva samples. The result is visible through the use of a handheld ultraviolet light-emitting diode flashlight. When the flashlight is used on the test strip, fluorescence is emitted and visible to the naked eye. The limit of detection for the method ranges from 1 to 10 gametocytes per μL of blood [34]. The test can be completed between 3 and 30 min. The approximate sensitivity of the method in symptomatic patients was 83% (95% CI, 61 to 95) when compared to PCR as the gold standard [34-38].

Advantages and Limitations the saliva-based diagnosis is noninvasive and the p LDH method employed is similar to that of blood-based RDTs. The sensitivity varies and test results need to be confirmed using microscopy. The limit of detection is also too high, and more research needs to be carried out to improve it to an acceptable range. It has the potential to be used not only for the diagnosis of symptomatic patients but also for asymptomatic carriers. This may be especially advantageous for control programs to identify and treat carriers of the parasite contributing to malaria transmission. The sensitivity of the test is comparable to that of current RDTs. Although the test is not yet commercialized, it can be estimated that the cost will be comparable to that of current blood based RDTs. Like blood based RDTs the test is not quantitative and will not be useful for determining parasitemia percentage in patients.

Urine-based malaria test

Urine malaria tests involve the detection of Plasmodium protein pHRP-2. A commercially available test, known as the Urine Malaria Test (UMT), involves dipping the test strip into a urine sample for two minutes, followed by incubation for twenty minutes [39]. Just like the commonly used RDT, a positive result is indicated by dark-colored lines on the test strip. However, the differentiating factor is that this method does not involve a finger stick and can be carried out non-invasively. Multicenter clinical trials conducted in Nigeria with febrile and afebrile patients, using the UMT, have confirmed the sensitivity and specificity of the method to be 79% and 89%, respectively when compared to the BinaxNOW Malaria Test kit. The sensitivity and specificity increased to 93% and 83%, respectively when used among febrile children under five years old [35]. The

limit of detection for this test is 125 parasites/ μL , and each test costs about \$1.50 [13,36]. Advantages and Limitations The urine-based malaria test is relatively affordable and does not require expensive equipment or highly trained personnel. The limitation of this test is that it only detects pHRP-2 from *P. falciparum* parasites.

Transdermal hemozoin detection

This method involves the detection of hemozoin-generated vapor nanobubbles using an ultrasound sensor. Hemozoin is the by-product of hemoglobin digestion by blood-stage malaria parasites. A short laser pulse administered to blood vessels through the skin localizes heat and evaporates the liquid around the hemozoin crystals. This liquid evaporation creates expanding and collapsing small-sized vapor nanobubbles inside the malaria parasite [37]. After the laser is activated, the probe can detect an acoustic pulse and generates an electrical signal as an acoustic trace. A patient with confirmed malaria was tested using the prototype, and hemozoin-generated vapor nanobubbles were detected. The short laser pulse administered is skin-safe and the test is estimated to take only a few seconds to perform and can detect 0.00034% parasitemia [38]. Since hemozoin clears from the blood within nine days, compared to pHRP-2 which is cleared only after several months, this method is expected to produce few false-positive results [40].

Advantages and Limitations The main advantages of this method are that it is non-invasive, requires no reagents and the results can be obtained within seconds. The method has a very low limit of detection and may potentially detect subclinical carriers which is helpful for disease surveillance. Highly trained personnel will be needed for the initial deployment of the method in both clinical and field settings. Researchers estimated that a battery-powered device will cost about US \$15,000, however, the cost of a single test could be lower than that of RDTs because a single device can be used for a larger number of patients [38]. Skin color has also been found to impact the results of the test and more studies are needed to address this issue. In addition, there is a need for further studies to determine the safety of extended laser application, even at low pulses, to the skin.

Conclusions and Recommendation

Although current diagnostic methods in use are not all perfect, they continue to play important roles in dealing with the current global malaria situation and decreasing the incidence of malaria. Numerous innovations continue in developing additional invasive as well as noninvasive and specific methods. Diagnostic tools are critical in ensuring that each patient receives the appropriate care. Currently used methods such as microscopy, RDTs, and PCR are not being utilized to their full capacity due to several barriers and limitations, such as cost, trained personnel, access to equipment, and unreliable electricity. In addition, not every febrile person is seen by a clinician, and they usually rely on self-medication using antimalarial drugs purchased from local pharmacies or retailers without a proper diagnosis and a prescription. This overuse and over-prescription of drugs are likely contributing to the emergence

of anti-malarial-resistant Plasmodium parasites. The “test and treat” strategy needs to become the norm everywhere to impede the emergence of drug resistance. RDTs, although not perfect, have the potential to result in a reduction of malaria incidence in endemic areas due to their ease of use and low cost. As disease incidence decreases and asymptomatic infections become more prevalent, RDTs will no longer be useful in tracking submicroscopic carriers due to their limited detection. Therefore, there will have to be a shift to implementing more sensitive and specific methods with lower limits of detection to diagnose and treat asymptomatic carriers. Interventions, such as providing malaria posts to every community to enable the diagnosis of febrile patients by a community health worker, have been shown to reduce malaria incidence if carried out properly. Malaria cases need to decline to reduce under-five and maternal mortality and achieve the United Nation’s sustainable development goal. It will be important to build trust between communities and health workers and to ensure hospitals and pharmacy retailers adherence to recommendations based on test results. This will require a multi-sectoral approach and political will to advocate for the implementation of control and elimination strategies. Further considerations are warranted for placing RDTs in pharmacies, drug shops, or malaria posts on one hand, and to educate people in seeking prompt treatment rather than alternative remedies to avoid hospital-based treatment on the other. Finally, the control of malaria continues to be challenged by the ubiquitous nature of the mosquito vector, widespread insecticide resistance, ever-threatening anti-malarial drug resistance, and the lack of an effective vaccine. Sensitive and accurate diagnosis and prompt treatment will continue to play a critical role in controlling malaria while other approaches become available to meet the challenge of malaria elimination. Other key requirements to achieve the goal of a malaria-free world will also rely on: (1) the recognition of trained community health workers paid fair wages as recognized members of the health system, (2) the establishment of malaria posts in endemic communities to provide access to diagnostics and treatment for those in hard-to-reach locations, (3) making RDTs available in local pharmacies and retail shops in resource-limited settings, (4) integrating malaria control programs into other disease programs of an individual country, and (5) political will for the long-term commitment of national financial resources [41-48].

References

1. WHO World Malaria Report 2019.
2. WHO Countries and Territories Certified Malaria-Free by WHO.
3. Mahmoudi S, Keshavarz H (2017) Efficacy of phase 3 trial of RTS, S/AS01 malaria vaccine: The need for an alternative development plan. Hum Vaccines Immunother 13: 2098-2101.
4. Ashley EA, Phyo AP, Woodrow CJ (2018) Malaria. Lancet 391: 21-27.
5. Singh B, Daneshvar C (2013) Human infections and detection of Plasmodium knowlesi. Clin Microbiol Rev 26(2): 165-184.
6. WHO Guidelines for the Treatment of Malaria.
7. White NJ, Pukrittayakamee S, Hien TT, Faiz MA, Mokuolu OA (2014) Malaria. Lancet 383: 723-735.
8. WHO Global Technical Strategy for Malaria 2016-2030.

9. Feachem RG, Chen I, Akbari O, Bertozzi V, Bhatt S, et al. (2019) Malaria eradication within a generation. *Lancet Comm* 394 (10203): 1056-1112.
10. Mathison B, Pritt B (2017) Update on malaria diagnostics and test utilization. *J Clin Microbiol* 55(7):2009-2017.
11. Cordray MS, Richards K (2012) Emerging nucleic acid-based tests for point-of-care detection of malaria. *Am J Trop Med Hyg* 87(2): 223-230.
12. Kolluri N, Klapperich CM, Cabodi M (2018) Towards lab-on-a-chip diagnostics for malaria elimination. *Lab Chip* 18(1): 75-94.
13. Pham NM, Karlen W, Beck HP, Delamarche E (2018) Malaria and the 'last' parasite: How can technology help? *Malar J* 17: 260.
14. Wilson M (2012) Malaria rapid diagnostic tests. *Clin Infect Dis* 54(11): 637-1641.
15. Amir A, Cheong FW, De Silva JR, Lau YL (2018) Diagnostic tools in childhood malaria. *Parasites Vectors* 11: 53.
16. Mouatcho JC, Goldring JP (2013) Malaria rapid diagnostic tests: Challenges and prospects. *J Med Microbiol* 62(10): 1491-1505.
17. Mukkala AN, Kwan J, Lau R, Harris D, Kain D (2018) An update on malaria rapid diagnostic tests. *Curr Infect Dis Rep* 20(12): 49.
18. WHO Malaria Rapid Diagnostic Test Performance: Summary Results of WHO Product Testing of Malaria RDTs: Round 1-7 (2008-2016).
19. Cunningham J, Jones S, Gatton ML, Barnwell JW, Cheng Q, et al. (2019) A review of the WHO malaria rapid diagnostic test product testing programme (2008-2018): Performance, procurement and policy. *Malar J* 18: 387.
20. Abba K, Deeks JJ, Olliaro P, Naing CM, Jackson SM, et al. (2011) Rapid diagnostic tests for diagnosing uncomplicated *P. falciparum* malaria in endemic countries. *Cochrane Database Syst Rev* 7: CD008122.
21. Oriero E, Jacobs J, Van GJ, Nwakanma D, Alessandro U (2015) Molecular-based isothermal tests for field diagnosis of malaria and their potential contribution to malaria elimination. *J Antimicrob Chemother* 70(1): 2-13.
22. Vasoo S, Pritt B (2013) Molecular Diagnostics and Parasitic Disease. *Clin Lab Med* 33(3): 461-503.
23. Roth J, Korevaar D, Leeflang M, Mens P (2015) Molecular malaria diagnostics: A systematic review and meta-analysis. *Crit Rev Clin Lab Sci* 53(2): 87-105.
24. Ragavan KV, Kumar S, Swaraj S, Neethirajan S (2018) Advances in biosensors and optical assays for diagnosis and detection of malaria. *Biosens Bioelectron* 105: 188-210.
25. Notomi T, Okayama H, Masubuchi H, Yonekawa T, Watanabe K (2000) Loop-mediated isothermal amplification of DNA. *Nucleic Acids Res* 28(12): E63.
26. Abdul GR, Al MA, Karanis P (2012) Loop-Mediated Isothermal Amplification (LAMP) for malarial parasites of humans: Would it come to clinical reality as a point-of-care test? *Acta Trop* 122(3): 233-240.
27. Polley SD, Mori Y, Watson J, Perkins M, González IJ et al. (2010) Mitochondrial DNA targets increase sensitivity of malaria detection using loop-mediated isothermal amplification. *J Clin Microbiol* 48(8): 2866-2871.
28. Li Y, Kumar N, Gopalakrishnan A, Ginocchio C, Manji R, et al. (2013) Detection and species identification of malaria parasites by isothermal t HDA amplification directly from human blood without sample preparation. *J Mol Diagn* 15(5): 634-641.
29. Mfuh KO, Yunga ST, Esemu LF, Bekindaka ON, Yonga J, et al. (2017) Detection of *Plasmodium falciparum* DNA in saliva samples stored at room temperature: Potential for a non-invasive saliva-based diagnostic test for malaria. *Malar J* 16(1): 434.
30. Singh R, Singh DP, Gupta R, Savargaonkar D, Singh OP (2014) Comparison of three PCR-based assays for the non-invasive diagnosis of malaria: Detection of *Plasmodium* parasites in blood and saliva. *Eur J Clin Microbiol Infect Dis* 33(9): 1631-1639.
31. Ongagna SY, Corstjens P, Geva E, Abrams WR, Barber CA (2013) Improved assay to detect *Plasmodium falciparum* using an uninterrupted, semi-nested PCR and quantitative lateral flow analysis. *Malar J* 12: 74.
32. Gbotosho GO, Happi CT, Folarin O, Keyamo O, Sowunmi A, et al. (2010) Multicenter pivotal clinical trial of urine malaria test for rapid diagnosis of *Plasmodium falciparum* malaria in children with acute uncomplicated malaria. *Am J Trop Med Hyg* 83(3): 496-501.
33. Ouattara A, Doumbo S, Saye R, Beavogui AH, Traoré B, et al. (2011) Use of a p LDH-based dipstick in the diagnostic and therapeutic follow-up of malaria patients in Mali. *Malar J* 10: 345.
34. Tao D, Mc Gill B, Hamerly T, Kobayashi T, Khare P, et al. (2019) A saliva-based rapid test to quantify the infectious subclinical malaria parasite reservoir. *Sci Transl Med* 11(473): eaan4479.
35. Oyibo WA, Ezeigwe N, Ntadom G, Oladosu OO, Rainwater LK, et al. (2016) Multicenter pivotal clinical trial of urine malaria test for rapid diagnosis of *Plasmodium falciparum* malaria. *J Clin Microbiol* 55(1): 253-263.
36. Fyodor Biotechnologies.
37. Lukianova HY, Campbell KM, Constantinou PE, Braam J, Olson JS, et al. (2014) Hemozoin-generated vapor nanobubbles for transdermal reagent- and needle-free detection of malaria. *Proc Natl Acad Sci* 111(3): 900-905.
38. Lukianova HE, Bezek S, Szigeti R, Khodarev A, Kelley T (2015) Transdermal diagnosis of malaria using vapor nanobubbles. *Emerg Infect Dis* 21(7): 1122-1127.
39. Oguonu T, Shu E, Ezeonwu BU, Lige B, Derrick A, et al. (2014) The performance evaluation of a Urine Malaria Test (UMT) kit for the diagnosis of malaria in individuals with fever in south-east Nigeria: Cross-sectional analytical study. *Malar J* 13: 403.
40. Grimberg BT, Grimberg KO (2016) Hemozoin detection may provide an inexpensive, sensitive, 1-minute malaria test that could revolutionize malaria screening. *Expert Rev Anti Infect Ther* 14(10): 879-883.
41. Britton S, Cheng Q, McCarthy JS (2016) Novel molecular diagnostic tools for malaria elimination: A review of options from the point of view of high-throughput and applicability in resource limited settings. *Malar J* 15: 88.
42. Chen I, Clarke SE, Gosling R, Hamainza B, Killeen G, et al. (2016) Asymptomatic malaria: A chronic and debilitating infection that should be treated. *PLoS Med* 13: e1001942.
43. Galatas B, Bassat Q, Mayor A (2016) Malaria Parasites in the Asymptomatic: Looking for the Hay in the Haystack. *Trends Parasitol* 32(4): 296-308.
44. Bousema T, Okell L, Felger I, Drakeley C (2014) Asymptomatic malaria infections: Detectability, transmissibility and public health relevance. *Nat Rev Microbiol* 12(12): 833-840.
45. Aguilar R, Machevo S, Menéndez C, Bardají A, Nhabomba A, et al. (2012) Comparison of placental blood microscopy and the ICT HRP2 rapid diagnostic test to detect placental malaria. *Trans R Soc Trop Med Hyg* 106(9): 573-575.
46. Fried M, Muehlenbachs A, Duffy PE (2012) Diagnosing malaria in pregnancy: An update. *Expert Rev. Anti Infect Ther* 10(10): 1177-1187.
47. WHO Malaria Elimination Certification Process.
48. WHO The E-2020 Initiative of 21 Malaria-Eliminating Countries: 2019 Progress Report.